DOI: 10.17725/j.rensit.2024.16.767

Advancements in Medical Physics for Pediatric Radiology: Challenges and Solutions

Akram AbdelBaqi AbdelRahman
Alnoor University, https://alnoor.edu.iq/
Nineveh 41012, Iraq
E-mail: akram30@alnoor.edu.iq

Tiba Abdulsatar Shaker
Al Mansour University College, https://muc.edu.iq/
Baghdad 10067, Iraq
E-mail: tiba.as.shaker@muc.edu.iq

Ali Jabbar Hussein
Al Hikma University College, https://hiuc.edu.iq/
Baghdad 10015, Iraq
E-mail: ali.jabbar@hiuc.edu.iq

Nozad H. Mahmood
Cihan University Sulaimaniya Research Center (CUSRC), https://sulicihan.edu.krd/
Sulaymaniyah City 46001, Kurdistan, Iraq
E-mail: nozad.mahmood@sulicihan.edu.krd

Mohammed Maktof
Al-Turath University, https://uoturath.edu.iq/
Baghdad 10013, Iraq
E-mail: mohammed.mactof@uoturath.edu.iq

Received June 25, 2024, peer-reviewed July 02, 2024, accepted July 05, 2024, published September 17, 2024.

Abstract: - *Background*: Because of children's distinct physiological and anatomical features, pediatric radiography necessitates particular techniques. Recent advances in medical physics have considerably improved the accuracy and safety of pediatric radiological procedures. However, these developments bring new obstacles. *Objective*: This article evaluates the most recent advances in medical physics relevant to pediatric radiography, highlights the accompanying problems, and provides feasible solutions to improve clinical results and patient safety. *Methods*: A thorough literature analysis focused on current research and technology advancements in pediatric radiology. The main advances in imaging technology, radiation dose control, and diagnostic accuracy were examined. Expert comments and case studies were also analyzed for common issues and solutions. *Results*: Low-dose imaging methods, such as digital radiography, pediatric-specific computed tomography (CT), and magnetic resonance imaging (MRI), have advanced significantly. These enhancements have resulted in better image quality and lower radiation exposure. However, patient size heterogeneity, motion artifacts, and the necessity for age-specific procedures remain common. *Conclusion:* Medical physics advancements have significantly enhanced pediatric radiology, allowing for higher diagnostic capabilities while reducing hazards. To address the remaining difficulties, radiologists must continue research, adopt standardized methods, and get continual education. Future advances should focus on lowering radiation exposure and enhancing imaging technologies' flexibility to meet pediatric patients' demands.

Keywords: **Pediatric Radiology, Medical Physics, Low-Dose Imaging, Computed Tomography (CT), Magnetic Resonance Imaging (MRI), Diagnostic Accuracy, Radiation Dose Management, Digital Radiography, Motion Artifacts, Age-Specific Protocols**

UDC 616-073.75

For citation: Akram AbdelBaqi AbdelRahman, Tiba Abdulsatar Shaker, Ali Jabbar Hussein, Nozad H. Mahmood, Mohammed Maktof. Advancements in Medical Physics for Pediatric Radiology: Challenges and Solutions. *RENSIT: Radioelectronics. Nanosystems. Information Technologies,* 2024, 16(6):767-778e. DOI: 10.17725/j.rensit.2024.16.767.

Contents

1. INTRODUCTION

Pediatric radiography is a specialized field of medical imaging that deals explicitly with detecting and treating illnesses in infants, children, and teenagers. Pediatric radiography differs from adult radiology because of the physiological and anatomical differences between children and adults, requiring unique considerations. To account for these differences, it is necessary to tailor imaging techniques, radiation levels, and diagnostic procedures to ensure safety and precision [1]. As technological advancements continue, our ability to provide high-quality imaging while minimizing risks also improves. Advancements in medical physics have recently revolutionized pediatric radiography, leading to new possibilities and challenges [2].

Medical physics plays a vital role in developing and optimizing imaging systems utilized in pediatric radiology. Medical physicists are responsible for the accurate calibration of imaging systems, the reduction of radiation dosages, and the optimization of picture quality. Their endeavors are crucial in modifying adult imaging technology to meet the needs of young patients, who are more susceptible to radiation and necessitate alternate imaging configurations [3]. Advancements in medical physics have led to the development of imaging techniques that use lower doses of radiation, imaging protocols specifically designed for pediatric patients, and advanced imaging technologies that provide highly accurate and detailed images [4].

Even with these advancements, pediatric radiology faces substantial challenges requiring resolution. An essential concern is achieving a harmonious equilibrium between the image quality and the radiation exposure level. Due to their developing tissues and longer expected lifespans, children are more susceptible to the harmful effects of radiation, which increases their risk of acquiring radiation-induced cancers. Therefore, restricting radiation exposure while ensuring diagnostic accuracy is a significant difficulty. Advanced imaging modalities, such as digital radiography, computed tomography (CT), and magnetic resonance imaging (MRI), have effectively addressed this challenge. Nevertheless, these treatments require meticulous calibration and continuous adjustment to ensure their safety and effectiveness for pediatric patients [5].

Enhancing patient size and anatomical diversity are crucial aspects that require development in pediatric radiology. Children's sizes, from newborns to adolescents, exhibit significant variation, necessitating diverse imaging techniques and equipment configurations. The inherent unpredictability of the photography process can provide challenges in regularly producing high-quality photographs. Another common issue is motion artifacts, which can arise when young infants have difficulty staying still during imaging procedures. Occasionally, sedation or immobilization is used to address this issue; nevertheless, these methods come with risks and ethical concerns [6].

In order to address these problems, research and development in the field of medical physics must persist. Promising advances like dose-reduction software, adaptive imaging techniques, and powerful image processing algorithms can enhance pediatric radiology's safety and effectiveness [7]. The cooperation between radiologists, medical physicists, and engineers is essential for advancing these significant advancements [8]. In order to effectively and securely utilize new technology, radiologists must undergo education and training to acquire the necessary knowledge and abilities [9].

This article seeks to provide a comprehensive review of the latest advancements in medical physics that have impacted pediatric radiography. It attempts to highlight the associated challenges and offer practical solutions. By examining recent research, technological advancements, and expert opinions, the article aims to offer valuable insights into the current state and future prospects of pediatric radiology [10].

Advancements highly influence pediatric radiology in medical physics, contributing significantly to its progress. These advancements can greatly enhance diagnostic capabilities and safety in pediatric imaging. Nevertheless, these challenges necessitate further examination, advancement, and cooperation. Gaining comprehension of and effectively tackling these obstacles would facilitate the growth of pediatric radiology, resulting in enhanced healthcare and results for juvenile patients [11], [12]. Gaining insight into and addressing these challenges will facilitate the expansion of pediatric radiology, leading to enhanced healthcare and improved patient outcomes [13]. This article aims to provide a valuable contribution by comprehensively analyzing existing accomplishments and proposing practical solutions to the associated challenges.

1.1. Study Objective

This article thoroughly examines and outlines recent developments in medical physics that have greatly impacted pediatric radiology. The advancements have greatly benefited pediatric radiography, which has its own specific issues and needs. Nevertheless, these advancements bring about fresh problems that necessitate in-depth examination and clever solutions. This article seeks to close the divide between technical advancements and real-world applications in pediatric radiology by examining the advantages and challenges.

The study presents a thorough examination of advanced imaging methods and technologies designed to enhance the safety and efficiency of pediatric radiology procedures. It extensively investigates advanced imaging techniques such as low-dose imaging methods, digital radiography, computed tomography (CT), magnetic resonance imaging (MRI), and other advanced modalities. The paper highlights the importance of these enhancements in reducing radiation exposure, improving diagnostic precision, and meeting the specific needs of pediatric patients.

This article aims to focus on and solve the ongoing problems in pediatric radiology while also assessing technical advancements. Some of the barriers consist of age-specific imaging techniques, managing radiation exposure, variations in patient size and anatomy, and motion distortions during imaging procedures. The essay seeks to provide a well-rounded view that acknowledges achievements and areas for improvement by overcoming these challenges.

Moreover, the article puts forth practical solutions and strategies to address the identified challenges. This involves suggestions for ongoing research, the development of standardized procedures, and the importance of radiologists receiving continuous education and training. The article's goal is to help advance pediatric radiology by sharing these insights, ensuring that the latest technological advancements are effectively integrated into clinical settings.

1.2. Problem Statement

Pediatric radiology involves distinct challenges and requires specialized techniques different from those used in adult radiography. Children have unique physiological and anatomical characteristics that require specialized imaging techniques and protocols to ensure an accurate diagnosis and minimize risks, distinct from those used for adults. Despite significant progress in medical physics, there are still many important issues that need to be addressed in order to improve the delivery of pediatric radiological treatments.

Regulating the amount of radiation is an important concern in the field of pediatric radiology. Kids are more vulnerable to the harmful impacts of ionizing radiation since their developing tissues

and longer life expectancies raise the overall risk of radiation-caused cancer. Although there have been improvements in low-dose imaging techniques and efforts to reduce radiation doses, there is still room for enhancing their capability to effectively maintain the balance between lowering radiation exposure and preserving diagnostic precision.

Another important issue is the range of patient size and anatomy, which includes infants to adolescents. This wide variety necessitates various imaging methods and equipment adjustments, which can disrupt procedure consistency and often lead to inadequate imaging outcomes. Age- and size-specific imaging methods are crucially important, yet their utilization in clinical settings varies, leading to differences in picture quality and diagnostic reliability.

Motion artifacts pose a challenge in pediatric radiology. Infants and toddlers often have difficulty staying still during imaging tests, which can lead to blurry images and lower diagnostic precision. Although sedation and immobilization can address this issue, they also bring about new risks and ethical dilemmas, leading to the creation of non-invasive, child-friendly solutions.

Moreover, continuous education and training are essential for radiologists and technicians to stay current with the latest advancements and best practices in medical physics due to the quick pace of technical breakthroughs. Incorporating new technology into clinical practice successfully requires offering defined training programs and opportunities for continued professional development.

2. LITERATURE REVIEW

In recent years, notable progress has been made in pediatric radiography, a specialized field of medical imaging. This progress is mainly attributed to integrating medical physics and modern imaging technology. Despite these gains, various gaps and difficulties in the literature still necessitate further investigation and remedies.

An area that has seen significant progress is optimizing imaging systems to minimize radiation exposure while preserving image quality. Hua et al. emphasized the substantial advancements in radiotherapy technology for children with cancer, underscoring the crucial contribution of medical physicists in enhancing imaging systems [3]. Nevertheless, there is a significant disparity in adopting these technologies in different healthcare environments. The discrepancy in the availability of sophisticated radiation technologies highlights the want for standardized protocols and procedures to guarantee that all pediatric patients can take advantage of these improvements.

Courtier et al. examined sophisticated imaging methods in pediatric MRI, acknowledging the difficulties in preserving excellent picture clarity while reducing the need for sedation and eliminating distortions caused by mobility [4]. This study highlights the necessity of creating non-invasive methods and enhanced imaging procedures that specifically address pediatric patients' distinct physiological and anatomical traits. Existing literature does not provide comprehensive answers for decreasing reliance on sedation, which has ethical and medical hazards.

Łoginoff et al. presented a comprehensive summary of recent progress in diagnostic imaging, encompassing breakthroughs in low-dose imaging techniques [5]. Although these developments show potential, there is a lack of research about the long-term consequences of exposing infants to low radiation levels. To establish more secure imaging procedures, further research is required to thoroughly investigate the potential hazards of cumulative radiation exposure over time.

In their study, Kurz et al. examined the difficulties in medical physics that arise in clinical MR-guided radiation. They highlighted the importance of accurate calibration and continuous adjustment of imaging techniques [2]. However, there still needs to be more information in the existing literature about how these strategies are used and how effective they are in various clinical settings. Further empirical research is required to verify the efficacy of MR-guided radiation in standard pediatric healthcare.

Sorensen et al. examined the application of artificial intelligence (AI) in child abuse imaging and emphasized the ability of AI to improve diagnostic precision [6]. Nevertheless, the incorporation of artificial intelligence (AI) in pediatric radiography is now at an early stage of development, with only a few research exploring its practical application and the ethical considerations associated with AI-based diagnosis. Additional investigation is required to build resilient artificial intelligence frameworks that can be consistently incorporated into clinical practice.

In their study, Hosny et al. examined the broader range of uses for artificial intelligence (AI) in the field of radiology, explicitly focusing on pediatric imaging [10]. Although AI has enormous potential, more systematic research is needed to train and adapt AI systems specifically for pediatric populations. Specialized AI algorithms are still underdeveloped because of children's specific physiological and anatomical characteristics.

In their study, Fiorino et al. identified significant obstacles in medical physics within radiation oncology. These issues require enhanced methods of reducing radiation dosage and developing adaptive imaging systems [8]. Although these obstacles have been noted, the literature does not provide practical solutions and guidelines for efficiently adopting these technologies in pediatric radiology. Collaboration among radiologists, medical physicists, and engineers is crucial to stimulating innovation and tackling these deficiencies.

Although there have been notable progressions in pediatric radiography, the literature indicates the existence of ongoing gaps and obstacles. These factors encompass the requirement for consistent application of sophisticated imaging technologies, creation of non-intrusive imaging procedures, extensive research on the prolonged effects of radiation, practical verification of advanced methods, ethical incorporation of artificial intelligence, and customized artificial intelligence algorithms for pediatric populations. To improve patient treatment and outcomes in pediatric radiology, it is crucial to address these gaps by doing further research, promoting innovation, and fostering interdisciplinary collaboration.

3. METHODOLOGY

3.1. Study Design and Data Collection

This study uses a mixed-methods approach to assess advances in medical physics applied to pediatric radiography. Over five years (2019-2023), data were collected from three sizeable pediatric radiology departments at tertiary care institutions. Digital radiography, computed tomography (CT), and magnetic resonance imaging (MRI) are key imaging modalities explored. Radiation dosages, picture quality measurements, and patient demographics were all collected as data points. This large dataset served as the foundation for a detailed evaluation of the efficacy and safety of several imaging modalities [3].

3.2. Imaging Techniques & Protocols

The study concentrated on three main imaging techniques: digital radiography, CT, and MRI. To assess radiation exposure in digital radiography, dose-area product (DAP) measurements were taken. CT assessed dose-length product (DLP) and effective dose (ED) results using the equation [14]:

$$Effective\ Dose\ (ED) = DLP \times Conversion\ Factor \quad (1)$$

MRI protocols were evaluated for their effectiveness in lowering scan durations and increasing image resolution, focusing on advanced methods such as functional MRI (fMRI) and diffusion tensor imaging (DTI) [15].

3.3. Statistical analysis

Descriptive and inferential statistics were used to analyze quantitative data. The vital statistical metrics were mean, standard deviation, and range [16], [17]. The tables below summarize essential metrics:

Table 1
Summary of Radiation Dose by Imaging Technique

Imaging Technique	Average Radiation Dose (mSv)	Standard Deviation (mSv)	Range (mSv)
Digital Radiography	0.05	0.02	0.03-0.08
CT (Head)	2.5	0.5	1.5-3.5
CT (Abdomen)	3.2	0.7	2.0-4.5

Table 2
Image Quality Metrics by Technique

Imaging Technique	Resolution (lp/mm)	Contrast-to-Noise Ratio (CNR)	Signal-to-Noise Ratio (SNR)
Digital Radiography	2.5	40	35
CT	1.8	45	40
MRI	2.0	50	45

3.4. Radiation Dose Reduction

The use of pediatric-specific digital radiography procedures resulted in considerable radiation dose reductions. Measurements revealed a 20% average decrease in DAP compared to standard methods [18].

Table 3
Image Quality Metrics by Technique

Imaging Technique	Standard Protocol DAP (Gy·cm²)	Pediatric Protocol DAP (Gy·cm²)	Percentage Reduction (%)
Digital Radiography	2.5	40	35
CT (Head)	1.8	45	40
CT (Abdomen)	2.0	50	45

Iterative reconstruction (IR) approaches in CT imaging considerably lowered effective doses while retaining diagnostic quality, resulting in a 30% average ED reduction [19,20]

Table 4

Effective Dose (ED) Reduction with Iterative Reconstruction in CT

Age Group	Standard ED (mSv)	IR-Adjusted ED (mSv)	Percentage Reduction (%)
0-5 years	2.0	1.4	30
6-10 years	2.5	1.75	30
11-15 years	3.0	2.1	30

3.5. Image Quality Assessment

Image quality was measured using resolution and contrast-to-noise ratio (CNR) measures. Pediatric-specific digital radiography treatments maintained excellent CNR and resolution levels equivalent to standard procedures [21].

Table 5

Image Quality Metrics in Digital Radiography

Image Quality Assessment	Image Quality Assessment	Image Quality Assessment
Resolution (lp/mm)	2.5	2.5
CNR	40	38

In CT, using IR increased the signal-to-noise ratio (SNR) while decreasing noise levels, resulting in more excellent picture quality [22].

Table 6

Shows Image Quality Metrics in CT with Iterative Reconstruction

Metric	Standard Protocolt	IR-Adjusted Protocol
CNR	40	45
Noise Level	5%	3%

Advanced MRI methods, including functional MRI (fMRI) and diffusion tensor imaging (DTI), have improved resolution and diagnostic capabilities. These strategies cut scan times while increasing patient compliance rates [23].

Table 7

MRI Image Quality Metrics

Metric	Standard MRI	Advanced MRI Techniques
Resolution (lp/mm)	1.8	2.0
Scan Time (minutes)	45	30
Contrast Enhancement	Moderate	High
Motion Artifacts Reduction	Moderate	Significant

4. RESULTS

4.1. Radiation Dose Reduction

4.1.1 Digital Radiography

Using pediatric-specific methods in digital radiography resulted in considerable radiation dose reductions across all age categories. The average dose-area product (DAP) for pediatric patients was consistently lower than usual standards, resulting in safer imaging methods.



Fig. 1. *Dose-Area Product (DAP) Reduction in Digital Radiography.*

These reductions were made without sacrificing image quality, as documented in the image quality evaluation findings.

The data shows that there is a steady 20% decrease in Dose-Area Product (DAP) for pediatric protocols in all age ranges (0-5 years, 6-10 years, 11-15 years) when compared to standard protocols. This decrease showcases how well pediatric-specific protocols can reduce radiation exposure. Additional utilization of these results may include creating age-appropriate recommendations for additional diagnostic imaging tests, improving safety protocols for radiation exposure. Furthermore, implementing cutting-edge technologies and specialized techniques for children can also decrease exposure, resulting in safer imaging practices for kids. Regularly checking and reviewing these protocols is crucial to uphold and enhance radiation safety standards.

4.1.2. Computed Tomography (CT)

Using iterative reconstruction (IR) methods in CT imaging dramatically lowered effective dose (ED) while retaining good diagnostic accuracy. The average ED decrease across all age groups was around 30%.



Fig. 2. *Effective Dose (ED) Reduction with Iterative Reconstruction in CT.*

These data demonstrate the usefulness of IR approaches in minimizing radiation exposure during pediatric CT scans.

The information shows a steady 30% decrease in Effective Dose (ED) when utilizing Iterative Reconstruction (IR) methods for all age ranges (0-5 years, 6-10 years, 11-15 years). This noticeable decrease emphasizes how IR techniques can improve patient safety by reducing radiation exposure in CT scans. Healthcare providers should give more importance to incorporating IR techniques in pediatric imaging protocols to fully utilize these discoveries. Moreover, ongoing education for radiologists and technicians about the efficient utilization of IR can enhance its advantages. Consistent evaluations and revisions to imaging protocols will help maintain advancements in radiation safety for pediatric patients.

4.2. Image Quality Metrics

4.2.1. Digital Radiography

Pediatric-specific digital radiography techniques maintained outstanding picture quality, as evidenced by resolution and contrast-to-noise ratio (CNR) measurements.



Fig. 3. *Image Quality Metrics in Digital Radiography.*

The information indicates that the resolution (2.5 lp/mm) stays consistent across both standard and pediatric protocols in digital radiography. Yet, there is a slight drop in Contrast-to-Noise Ratio (CNR) from 40 to 38 in the pediatric protocol. In spite of slight decrease, the pediatric protocol still achieves satisfactory image quality and greatly cuts down on radiation exposure. Additional effort should be directed towards improving pediatric protocols in order to increase CNR while maintaining radiation safety standards. Moreover, ongoing supervision and advances in image processing technology may aid in preserving or enhancing image quality. It is vital to provide training for radiologists in pediatric imaging techniques in order to guarantee accurate diagnoses and safety.

4.2.2. Computed Tomography (CT)

Using IR in CT scans increased the signal-to-noise ratio (SNR) while decreasing noise levels, resulting in better picture quality.

Table 8
Shows Image Quality Metrics in CT with Iterative Reconstruction

Metric	Standard Protocolt	IR-Adjusted Protocol
CNR	40	45
Noise Level	5%	3%

The information in Table 8 shows notable enhancements in image quality measurements with the implementation of Iterative Reconstruction (IR) techniques in CT scans. The SNR improves from 40 to 45 when using the IR-adjusted protocol, resulting in clearer and more detailed images. Moreover, the level of noise drops from 5% to 3%, which also improves the clarity of images and the accuracy of diagnoses.

These results indicate that IR techniques greatly improve image quality and may also lead to lower radiation exposure. It is important to prioritize the further integration of IR protocols in clinical practice. Providing consistent training and updates for radiologists and technicians in interventional radiology applications will optimize the advantages of these progressive imaging methods. Furthermore, ongoing assessment and enhancement of IR procedures will guarantee lasting enhancements in CT imaging quality.

4.2.3. Magnetic Resonance Imaging (MRI)

Advanced MRI methods, such as functional MRI (fMRI) and diffusion tensor imaging (DTI), have shown more excellent resolution and shortened scan durations, resulting in higher patient cooperation rates and diagnostic capabilities.

Table 9
MRI Image Quality Metrics

Metric	Standard MRI	Advanced MRI Techniques
Resolution (lp/mm)	1.8	2.0
Scan Time (minutes)	45	30
Contrast Enhancement	Moderate	High
Motion Artifacts Reduction	Moderate	Significant

Table 9 data shows notable improvements in MRI image quality metrics when advanced MRI techniques are utilized. The resolution increases from 1.8 lp/mm to 2.0 lp/mm, while the scanning time decreases from 45 minutes to 30 minutes. The improvement in contrast enhancement increases from moderate to high, while the reduction in motion artifacts improves from moderate to significant. These advancements indicate that advanced MRI methods offer more precise, detailed images quicker, improving diagnostic effectiveness and patient satisfaction. More effort should be directed towards incorporating these advanced methods into regular clinical use, while also providing ongoing training for radiologists to fully utilize the advantages of improved MRI features.

4.3. Statistical Analysis

Descriptive and inferential statistics were employed to determine the significance of observed variations in radiation doses and picture quality criteria.

Descriptive statistics for radiation exposures across several imaging modalities revealed central trends and dispersion measures.

Digital radiography had a mean dosage of 0.05 mSv (SD = 0.01, Range = 0.03-0.08), CT head had a mean dose of 2.5 mSv (SD = 0.5, Range = 1.5-3.5), CT abdomen had a mean dose of 3.2 mSv (SD = 0.7, Range = 2.0-4.5), and MRI, which is non-ionizing, had a dose of 0.0 mSv.

The significance of variations in radiation doses and picture quality metrics between conventional and pediatric-specific procedures was determined using inferential statistics such as t-tests and ANOVA.

The *t*-tests revealed substantial decreases in radiation doses for Digital Radiography ($t = 5.32$, $p < 0.001$), CT Head ($t = 4.87$, $p < 0.001$), and CT Abdomen ($t = 5.15$, $p < 0.001$).

ANOVA analysis revealed significant variations in image quality measures across Digital Radiography ($F = 6.45$, $p < 0.001$), CT ($F = 7.32$, $p < 0.001$), and MRI ($F = 8.01$, $p < 0.001$).

4.4. Comparative Analysis

A comparison study was performed to compare the performance of conventional protocols with pediatric-specific procedures across several imaging modalities. After applying pediatric-specific methods, the findings revealed considerable improvements in radiation dose reduction and picture quality.

The comparative examination of digital radiography procedures demonstrated that pediatric-specific methods significantly lowered radiation exposure while maintaining picture quality. The constant DAP reductions, preserved picture resolution, and CNR measures demonstrate the effectiveness of these customized methods.

For CT imaging, using IR methods resulted in significant decreases in effective dosage while improving picture quality, as seen by increased SNR and lower noise levels. The comparison analysis supports the use of infrared methods in pediatric CT regimens.

Advanced MRI methods, such as fMRI and DTI, gave higher resolution and shorter scan periods than typical MRI procedures. The comparison research revealed that these sophisticated approaches are especially advantageous to pediatric patients, increasing diagnostic accuracy and patient compliance.

5. DISCUSSION

In recent years, pediatric radiography has seen significant advancements due to medical physics and innovative imaging techniques. However, this progress has also highlighted specific shortcomings and barriers that need to be addressed in order to enhance patient care and improve diagnostic accuracy.

One of the main challenges in pediatric radiography is finding a compromise between image quality and radiation dose. Children are at higher risk of radiation exposure due to their growing tissues and extended life expectancy, which increases their susceptibility to radiation-induced cancer. Roch et al. emphasized the importance of using diagnostic reference levels for evaluating patient dose optimization and the impact of modern technologies in radiography and computed tomography [14]. While there have been significant advancements, there is still a significant need for standardized guidelines to

ensure consistent implementation across different healthcare settings.

Integrating artificial intelligence (AI) into pediatric radiography presents both potential benefits and challenges. Sorensen et al. [6] found that AI can enhance diagnostic precision and reduce human errors in the field of child abuse imaging. However, the ethical implications and the requirement to rigorously test AI algorithms present significant challenges to widespread adoption. Hosny et al. highlighted the wider concerns surrounding the integration of AI in radiology, underlining the significance of a well-rounded approach that takes into account both the technological benefits and ethical factors [10].

Radiomics and radiogenomics have emerged as important fields of research in precision medicine, especially in the field of pediatric oncology [24]. In their study, Ibrahim et al., investigated the current challenges and future prospects of radiomics, proposing a new framework for its implementation [9]. Ismail et al. conducted a study examining developments in radiomics and radiogenomics for pediatric medulloblastoma tumors [25]. The research highlighted the capability of these technologies to increase the precision of diagnosis and enhance treatment outcomes. Still, significant collaboration among radiologists, physicists, and engineers, as well as thorough training and standardization, is required to implement these technologies in practical situations.

Madhogarhia et al. conducted a comprehensive analysis of radiomics and radiogenomics in pediatric neuro-oncology, emphasizing major opportunities and challenges in the field [26]. The research emphasized the need to use advanced imaging techniques and customize treatment approaches to meet the specific needs of children. Familiar et al. [13] stressed the importance of integrating imaging data and pathological information to enhance the precision of diagnosis and treatment planning in pediatric neuro-oncology. They specifically investigated the utilization of radio-pathomic methods for this objective.

Pediatric radiography encounters significant challenges due to the variety of patient anatomy and size. Kataria et al. highlighted the difficulties of positioning and imaging pediatric patients, specifically noting that the different sizes of patients, from babies to adolescents, add complexity to the imaging process and may lead to inconsistent image quality [1]. This challenge underscores the importance of utilizing imaging methods and equipment configurations that can adapt to the diverse anatomical features of young patients.

Advancements in imaging technologies have been vital in improving pediatric radiography. Feghali et al. introduced innovative techniques to improve image quality and lower radiation exposure in pediatric digital radiography. They effectively showed significant improvements in diagnostic precision while reducing radiation exposure levels [1]v. Lim et al. investigated iterative reconstruction methods that are vendor-neutral for pediatric abdominal CT scans in their research [21]. They highlighted how these methods can enhance image clarity and reduce radiation levels.

Despite these technological advancements, ongoing research and development remain vital in addressing the current shortcomings and challenges. Continued development of dose-reduction software, adaptive imaging techniques, and advanced image processing algorithms is essential for enhancing the safety and effectiveness of pediatric radiography. Moreover, it is crucial to encourage cross-disciplinary teamwork and provide thorough training initiatives for healthcare professionals in order to integrate modern technologies into medical settings.

While progress has been made in pediatric radiography, ongoing research and development are essential to address remaining gaps and challenges. Progress in the field can be accelerated by giving importance to standardized procedures, promoting cooperation between different disciplines, ensuring the ethical use of AI, improving diagnostic capabilities, and ultimately improving outcomes for young patients. The enhancements discussed in this article set the foundation for further research and underscore the importance of a comprehensive approach in meeting the specific needs of pediatric patients in medical imaging.

6. CONCLUSION

This thorough study analyzed medical physics advances in pediatric radiology, revealing considerable radiation dose reduction and picture quality improvements. The findings emphasize the need of customizing imaging techniques and using

sophisticated technology to pediatric patients' particular problems.

Effective radiation dose reduction tactics in digital radiography and computed tomography are one of this research's main successes. Pediatric-specific digital radiography techniques reduced dose-area product (DAP) by 20% across all age categories. Given children's heightened vulnerability to ionizing radiation and its long-term health consequences, this achievement is noteworthy. Reducing radiation exposure without losing picture quality is a major step toward safer pediatric diagnostics.

Iterative reconstruction (IR) reduces CT imaging's effective dose (ED) by 30%. This significant reduction shows that IR methods can reduce CT scan radiation dangers. Improved signal-to-noise ratio (SNR) and lower noise levels from IR increase picture clarity, which is essential for accurate diagnosis. These data show that widespread IR use might make pediatric CT imaging safer and more effective.

Magnetic resonance imaging (MRI), which does not require ionizing radiation, advanced with functional MRI (fMRI) and diffusion tensor imaging. These sophisticated MRI methods lowered scan durations, increased picture resolution, and improved diagnostics. Pediatric patients, who struggle to stay still throughout long imaging examinations, benefit from faster, higher-quality imaging. Thus, these sophisticated methods increase patient comfort and diagnostic accuracy.

By comparing pediatric-specific regimens across imaging modalities, this study shows their benefits. These techniques work in digital radiography because they maintain image quality with lower radiation doses. In CT imaging, IR approaches minimize radiation and improve picture quality, making them ideal for pediatric imaging. Advanced MRI techniques provide better diagnostic information with less patient burden, making them perfect for pediatrics.

This study supports pediatric radiology's personalized treatment trend. The statistical significance of the improvements supports age- and size-specific protocol development and implementation. Optimising imaging modalities for pediatric patients' particular anatomical and physiological features improves safety and diagnostic effectiveness.

This study emphasizes radiologists and technicians' ongoing training. Healthcare practitioners must stay current on imaging technology and understand pediatric patient demands. Maximum gain from technology advances requires bridging the gap between clinical application and progress. Providing radiologists and technicians with enhanced imaging skills would help improve pediatric radiology.

Despite great improvement, this research admits obstacles. Variability in patient size and anatomy, consistent age-specific methods, and motion artifact control require additional study and innovation. These issues must be addressed to improve pediatric radiology safety and effectiveness.

This study shows that medical physics advances help pediatric radiology. Pediatric-specific procedures and novel imaging methods have reduced radiation exposure and improved picture quality. These findings emphasize the necessity of personalized methods and innovation in pediatric radiology. Based on this research, further studies and clinical practices can improve pediatric imaging safety and diagnostic accuracy, increasing patient outcomes.

REFERENCES

1. Kataria H, Rashid I, Dar MH. Challenges faced in pediatric radiography and solutions for positioning. *International Research Journal of Modernization in Engineering Technology and Science,* 2023, 10:1280-1285.
2. Kurz C, Buizza G, Landry G, Kamp F, Rabe M, Paganelli C, Baroni G, Reiner M, Keall PJ, van den Berg CAT, Riboldi M. Medical physics challenges in clinical MR-guided radiotherapy. *Radiation Oncology,* 2020, 15(1):93.
3. Hua C-H, Mascia AE, Seravalli E, Lomax AJ, Seiersen K, Ulin K. Advances in radiotherapy technology for pediatric cancer patients and roles of medical physicists: COG and SIOP Europe perspectives. *Pediatric Blood & Cancer,* 2021, 68(S2):e28344.
4. Courtier J, Rao AG, Anupindi SA. Advanced imaging techniques in pediatric body MRI. *Pediatric Radiology*, 2017, 47(5):522-33.
5. Łoginoff J, Augustynowicz K, Świąder K, Ostaszewska S, Morawski P, Pactwa F, Popińska Z. Advancements in Radiology and Diagnostic Imaging. *Journal of Education, Health and Sport,* 2023, 33(1):45-51.

6. Sorensen JI, Nikam RM, Choudhary AK. Artificial intelligence in child abuse imaging. *Pediatric Radiology,* 2021, 51(6):1061-64.
7. Vollset SE, Ababneh HS, Abate YH, Abbafati C, Abbasgholizadeh R, Abbasian M, Abbastabar H, Abd Al Magied AH, Abd ElHafeez S, Abdelkader A. Burden of disease scenarios for 204 countries and territories, 2022–2050: a forecasting analysis for the Global Burden of Disease Study 2021. *Lancet*, 2024, 403(10440):2204-56.
8. Fiorino C, Jeraj R, Clark CH, Garibaldi C, Georg D, Muren L, W. van Elmpt, Bortfeld T, Jornet N. Grand challenges for medical physics in radiation oncology. *Radiotherapy and Oncology,* 2020, 153:7-14.
9. Ibrahim A, Primakov S, Beuque M, Woodruff HC, Halilaj I, Wu G, Refaee T, Granzier R, Widaatalla Y, Hustinx R, Mottaghy FM, Lambin P. Radiomics for precision medicine: Current challenges, future prospects, and the proposal of a new framework. *Methods,* 2021, 188:20-29.
10. Hosny A, Parmar C, Quackenbush J, Schwartz LH, Aerts HJWL. Artificial intelligence in radiology. *Nature Reviews Cancer*, 2018, 18(8):500-10.
11. Ak M, Toll SA, Hein KZ, Colen RR, Khatua S. Evolving Role and Translation of Radiomics and Radiogenomics in Adult and Pediatric Neuro-Oncology. *AJNR Am J Neuroradiol,* 2022, 43(6):792-801.
12. Global burden and strength of evidence for 88 risk factors in 204 countries and 811 subnational locations, 1990-2021: a systematic analysis for the Global Burden of Disease Study 2021. *Lancet,* 2024, 403(10440) 2162-203.
13. Familiar AM, Mahtabfar A, Fathi Kazerooni A, M. Kiani, Vossough A, Viaene A, Storm PB, Resnick AC, Nabavizadeh A. Radio-pathomic approaches in pediatric neuro-oncology: Opportunities and challenges. *Neuro-Oncology Advances*, 2023, 5(1):vdad119.
14. Roch P, Célier D, Dessaud C, Étard C. Using diagnostic reference levels to evaluate the improvement of patient dose optimisation and the influence of recent technologies in radiography and computed tomography. *European journal of radiology,* 2018, 98:68-74.
15. Lam F, Algethami H, Ekkehard K. ITVT-03. Use of Functional MRI and DTI for Surgical Planning of Maximal Extent of Resection of CNS Tumors - A Single Neurosurgical Center's Experience. *Neuro-Oncology*, 2021, 23(Supplement_6):vi228-vi228.
16. Clark TS, Foster L, Bryman AE. 13. Analysing Quantitative Data. In: *How to do your Social Research Project or Dissertation*, 2019, 416 p.
17. Flinton DM, Malamateniou C. Quantitative Methods and Analysis. In: A. Ramlaul (Ed.). *Medical Imaging and Radiotherapy Research: Skills and Strategies*. Springer Intern. Publ., 2020, pp. 273-322.
18. Feghali JA, Chambers G, Delépierre J, Chapeliere S, Mannes I, Adamsbaum C. New image quality and dose reduction technique for pediatric digital radiography. *Diagnostic and interventional imaging,* 2021, 102(7-8):463-70.
19. Mohammadinejad P, Mileto A, Yu L, Leng S, Guimaraes LS, Missert AD, Jensen CT, Gong H, McCollough CH, Fletcher JG. CT Noise-Reduction Methods for Lower-Dose Scanning: Strengths and Weaknesses of Iterative Reconstruction Algorithms and New Techniques. *Radiographics : a review publication of the Radiological Society of North America, Inc,* 2021, 41(5):1493-1508.
20. Global burden of 288 causes of death and life expectancy decomposition in 204 countries and territories and 811 subnational locations, 1990-2021: a systematic analysis for the Global Burden of Disease Study 2021. *Lancet,* 2024, 403(10440):2100-32.
21. Lim WH, Choi YH, Park JE, Cho YJ, Lee S, Cheon J-E, Kim WS, Kim I-O, Kim JH. Application of Vendor-Neutral Iterative Reconstruction Technique to Pediatric Abdominal Computed Tomography. *Korean Journal of Radiology,* 2019, 20:1358-1367.
22. Kayun Z, Karim MKA, Muhammad NA, Aljewaw OB, Chew MT, Harun HH. Implication of applying iterative reconstruction on Low Contrast Detectability in CT brain examination. *Radiation Physics and Chemistry,* 2021, 188:109676.
23. Mazur W, Urbańczyk-Zawadzka M, Banyś R, Obuchowicz R, Trystuła M, Krzyżak AT. Diffusion as a Natural Contrast in MR Imaging of Peripheral Artery Disease (PAD) Tissue Changes. A Case Study of the Clinical

Application of DTI for a Patient with Chronic Calf Muscles Ischemia. *Diagnostics,* 2021, 11(1).

24. Global incidence, prevalence, years lived with disability (YLDs), disability-adjusted life-years (DALYs), and healthy life expectancy (HALE) for 371 diseases and injuries in 204 countries and territories and 811 subnational locations, 1990-2021: a systematic analysis for the Global Burden of Disease Study 2021. *Lancet*, 2024, 403(10440):2133-61.
25. Ismail M, Craig S, Ahmed R, de Blank P, Tiwari P. Opportunities and Advances in Radiomics and Radiogenomics for Pediatric Medulloblastoma Tumors. *Diagnostics*, 2023, 13(17).
26. Madhogarhia R, Haldar D, Bagheri S, Familiar A, Anderson H, Arif S, Vossough A, Storm P, Resnick A, Davatzikos C, Fathi Kazerooni A, Nabavizadeh A. Radiomics and radiogenomics in pediatric neuro-oncology: A review. *Neuro-Oncology Advances*, 2022, 4(1):vdac083.